

Clinical recommendations regarding use of cone beam computed tomography in orthodontics. Position statement by the American Academy of Oral and Maxillofacial Radiology

American Academy of Oral and Maxillofacial Radiology

Aims. To summarize the potential benefits and risks of maxillofacial cone beam computed tomography (CBCT) use in orthodontic diagnosis, treatment and outcomes and to provide clinical guidance to dental practitioners.

Methods. This statement was developed by consensus agreement of a panel convened by the American Academy of Oral and Maxillofacial Radiology (AAOMR). The literature on the clinical efficacy of and radiation dose concepts associated with CBCT in all aspects of orthodontic practice was reviewed.

Results. The panel concluded that the use of CBCT in orthodontic treatment should be justified on an individual basis, based on clinical presentation. This statement provides general recommendations, specific use selection recommendations, optimization protocols, and radiation-dose, risk-assessment strategies for CBCT imaging in orthodontic diagnosis, treatment and outcomes.

Conclusions. The AAOMR supports the safe use of CBCT in dentistry. This position statement is periodically revised to reflect new evidence and, without reapproval, becomes invalid after 5 years. (Oral Surg Oral Med Oral Pathol Oral Radiol 2013;116:238-257)

Malocclusions and craniofacial anomalies adversely affect quality of life. Orthodontics and dentofacial orthopedic treatment address the correction of malocclusions and facial disproportions due to dental/skeletal discrepancies to provide esthetic, psychosocial, and functional improvements. For almost a century, two-dimensional (2D) planar radiographic imaging and cephalometry have been used to assess the interrelationships of the dentition, maxillofacial skeleton, and soft tissues in all phases of the management of orthodontic patients, including diagnosis, treatment planning, evaluation of growth and development, assessment of treatment progress and outcomes, and retention. However, the limitations of 2D imaging have been realized for decades as many orthodontic and dentofacial orthopedic problems involve the lateral or “third dimension.”¹⁻³ For instance, relapse of and unfavorable responses to orthodontic therapy remain poorly understood despite implications that considerations in the transverse plane are important factors in stability.⁴ For years, multiple radiographic projections were obtained to attempt to display complex anatomic relationships and surrounding structures; however, interpreting multiple-image inputs is challenging. With the increasing availability of multi-slice computed tomography (CT) and, more recently, cone beam computed tomography (CBCT), visualization of these relationships in three dimensions is now feasible.

SCOPE AND PURPOSE OF THE RECOMMENDATIONS AND CONCLUSIONS

This position statement was developed by board-certified orthodontists and oral and maxillofacial radiologists convened by the American Academy of Oral and Maxillofacial Radiology (AAOMR). Their objectives were to 1) review and evaluate critically the current science, guidance and other resources available from professional organizations on the clinical benefits and potential limitations of the use of CBCT in orthodontics, and 2) develop consensus derived, orthodontic-specific clinical guidelines. Imaging selection recommendations, optimization protocols and radiation-dose, risk-assessment strategies were developed to assist professional clinical judgment on the use of CBCT in orthodontics. The panel concluded that there is no clear evidence to support the routine use of ionizing radiation in standard orthodontic diagnosis and treatment planning, including the use of CBCT.

BACKGROUND

Imaging considerations in orthodontic therapy

One purpose of radiographic imaging in orthodontics is to supplement clinical diagnosis in the pretreatment assessment of the orthodontic patient. Radiographic imaging may also be performed during treatment to assess the effects of therapy and posttreatment to monitor stability and outcome. Imaging for a specific orthodontic patient occurs in at least three stages: 1) selection of the most appropriate radiographic imaging technique, 2) acquisition of appropriate images, and 3) interpretation of the images obtained. In some instances, these steps need to be repeated. Selection of the appropriate radiographic imaging technique (or techniques) is

Received for publication Jun 1, 2013; accepted for publication Jun 3, 2013.

© 2013 Published by Elsevier Inc.

2212-4403/\$ - see front matter

<http://dx.doi.org/10.1016/j.oooo.2013.06.002>

based on the principle that practitioners who use imaging with ionizing radiation have a professional responsibility of beneficence—that imaging is performed to “serve the patient’s best interests.” This requires that each radiation exposure is justified clinically and that procedures are applied that minimize patient radiation exposure while optimizing maximal diagnostic benefit. The extension of this principle, referred to as the “as low as reasonably achievable” (ALARA),⁵ to CBCT imaging is supported by the American Dental Association.⁶ Justification of every radiographic exposure must be based primarily on the individual patient’s presentation including considerations of the chief complaint, medical and dental history, and assessment of the physical status (as determined with a thorough clinical examination) and treatment goals.⁶

In 1987, a panel of representatives from general dentistry and various academic disciplines in the United States was convened by the Food and Drug Administration. This panel published broad selection recommendations for intraoral radiographic examinations.⁷ These were updated in 2004.^{8,9} The guidelines suggest that for monitoring growth and development of children and adolescents, “clinical judgment be used in determining the need for, and type of radiographic images necessary for, evaluation and/or monitoring of dentofacial growth and development.” In both the European Union¹⁰⁻¹² and the United Kingdom¹³ orthodontic imaging guidelines state that there is neither an indication for taking radiographs routinely before clinical examinations nor for taking a standard series of radiographic images for all orthodontic patients. The latter document provides clinical decision algorithms based on the ages of the patients (less than or over 9 years of age) and clinical presentation (delayed or ectopic eruption, crowding, or anteroposterior discrepancies such as overjet or overbite, etc.).

CBCT imaging in orthodontics

There has been a dramatic increase in the use of CBCT in dentistry over the last decade. This technology has found particular applications in orthodontics for diagnosis and treatment planning for both adult and pediatric patients.¹⁴⁻²⁰ CBCT imaging provides two unique features for orthodontic practice. The first is that numerous linear (e.g., lateral and posteroanterior cephalometric images) or curved planar projections (e.g., simulated panoramic images) currently used in orthodontic diagnosis, cephalometric analysis, and treatment planning can be derived from a single CBCT scan. This provides for greater clinical efficiency. The second, and most important, is that CBCT data can be reconstructed to provide unique images previously unavailable in orthodontic practice. Innately CBCT data are presented as inter-relational undistorted images in three orthogonal planes (i.e., axial, sagittal, and coronal); however,

software techniques are readily available (e.g., maximum intensity projection and surface or volumetric rendering) that provide three-dimensional visualization of the maxillofacial skeleton, airway space and soft tissue boundaries such as the facial outline. The current diagnostic uses of CBCT are summarized in [Appendix A](#).²¹⁻¹⁵⁸

Evidence based assessments

The potential for extracting additional diagnostic information from volumetric imaging and the technical ease of obtaining scans has led some clinicians and manufacturers to advocate the replacement of current conventional imaging modalities with CBCT for standard orthodontic diagnosis and treatment.^{15,18,159,160} Although CBCT imaging increases clinician confidence in orthodontic diagnosis¹⁶¹ and has demonstrated clinical efficacy in altering treatment planning for impacted maxillary canines,^{37,43,161} unerupted teeth, severe root resorption, and severe skeletal discrepancies,¹⁶¹ no benefit has been demonstrated for patients specifically referred for abnormalities of the temporomandibular joint, airway assessment or dental crowding.¹⁶¹ Despite the number of publications on the use of CBCT for specific orthodontic applications, most are observational studies of diagnostic performance and efficacy with wide ranging methodological soundness.¹⁶² Few authors have presented higher levels of evidence and measured the impact of CBCT on orthodontic diagnosis and treatment planning decisions.

Fundamentals to guideline development are systematic reviews of the published literature. Systematic reviews use well-defined and reproducible literature search strategies to identify evidence focused on a specific research question. Evidence is graded according to its level of methodological rigor (or quality), relevance and strength. There is a lack of CBCT-orthodontic systematic reviews. There is a need for rigorous investigation on the efficacy of CBCT imaging for all aspects of orthodontics related to its influence on therapy decisions and ultimately patient outcome.¹⁶³ Because of the lack of CBCT-orthodontic systematic reviews, the panel used consensus and published criteria.¹⁶⁴⁻¹⁶⁸ to develop three hierarchical recommendations for CBCT imaging in orthodontics ([Table I](#)). An important consideration in the use of CBCT is that ionizing radiation is a risk to patient health.

Radiation dose considerations in orthodontics

There are two broad potential harmful effects of ionizing radiation in orthodontics. The first is deterministic effects that cause the death of cells from high doses over short periods of time and usually occur only after thresholds are reached. Below these thresholds no clinical change has been reported. These levels are never reached for a single exposure in the diagnostic range

Table I. Panel consensus recommendations for use of CBCT imaging*

<i>Recommendation</i>	<i>Consensus level</i>	<i>Definition</i>
Likely indicated	I	The use of CBCT imaging is indicated in most circumstances for this clinical condition. There is an adequate body of evidence to indicate a favorable benefit from the procedure relative to the radiation risk in the majority of situations.
Possibly indicated	II	The use of CBCT imaging may be indicated in certain circumstances for this clinical condition. There is a sufficient body of evidence to indicate a possible favorable benefit from the procedure relative to the radiation risk in many situations.
Likely not indicated	III	The use of CBCT imaging is not indicated in the majority of circumstances for this clinical condition. There is an insufficient body of evidence to indicate a benefit from the procedure relative to the radiation risk in most situations.

*In the future, if CBCT imaging radiation levels are equivalent to conventional modalities, this table may be less relevant.

used in conventional oral and maxillofacial radiology. They do, however, occur in dental patients who have cancer and undergo radiotherapy to the head and neck region. One example of this is radiation-induced oral mucositis. The second effect is a stochastic effect that irreversibly alters the cells, usually by damaging cellular DNA. Such damage can result in cancer. The long-term risk associated with diagnostic radiographic imaging is radiation-induced carcinogenesis. Unlike deterministic effects, stochastic effects can result from low levels of radiation that are cumulative over time.

Assessment of the risks associated with the use of ionizing radiation for diagnostic imaging is an important **public health issue**. Recent reports have increased concerns over the potential association between radiation exposure and cancer. In one article, a relationship was found between intracranial meningiomas and dental radiographic procedures¹⁶⁹; however, numerous rebuttal articles have highlighted limitations in this study.¹⁷⁰⁻¹⁷³ Most recently, the results of a retrospective cohort study provide evidence of a link between exposure to radiation from medical CT and cancer risk

in children.¹⁷⁴ It was found that children and young adults who received radiation doses from the equivalent of 2 or 3 medical CT scans of the head have almost triple the risk of developing leukemia or brain cancer later in life. Medical CT head scans may have an effective dose of up to 2000 μSv¹⁷⁵; however, for CT examinations with dental protocols, substantial reductions to less than 1000 μSv have been reported.^{159,176-184} Most CBCT examinations impart a fraction of medical CT effective dose; however, doses vary considerably among CBCT units.^{90,137,159,176-196}

Low-dose radiographic procedures (including maxillofacial CBCT) are those that result in doses below about 1,00,000 μSv. The risk of cancer induction caused by low-dose radiographic procedures is difficult to assess. While there is lack of agreement among radiation epidemiologists and radiobiologists, there is consensus among the four authoritative agencies in the United States responsible for developing public-health, radiation-safety directives that for stochastic risks, such as carcinogenesis, the risks should be considered to be linearly related to doses, down to the lowest doses.¹⁹⁷⁻²⁰⁰

The assessment of risk is, however, confounded in that people are exposed to background radiation, including cosmic radiation from airline flights and/or living at high altitudes. For this position statement, the panel reviewed information on the potential health effects of exposure to diagnostic ionizing radiation. There is neither convincing evidence for carcinogenesis at the level of dental exposures, nor the absence of evidence of such damage. This situation is unlikely to change in the near future. In the absence of evidence of a threshold dose, it is prudent, from a patient-policy perspective, to assume that such a risk exists. This implies that there is no safe limit or “safety zone” for ionizing radiation exposure in diagnostic imaging. Every exposure cumulatively increases the risk of cancer induction. Consequently, to be cautious, the guidelines presented in this position statement are focused on minimizing or eliminating unnecessary radiation exposure in diagnostic imaging.

The overall biological effect of exposure to ionizing radiation, expressed as the risk of cancer development over a lifetime, is determined from absorbed radiation dose to specific organs in combination with weighting factors that account for differences in exposed-tissue sensitivity and patient susceptibility factors such as gender and age. For this position statement, the International Commission on Radiological Protection (ICRP)’s effective dose (*E*) method was used to estimate whole body dose and measure stochastic radiation risks to patients based on evidence of biological effects currently available.²⁰¹ Effective dose is calculated by multiplying organ doses by risk weighting factors (which are the organs’ relative radiosensitivities to developing cancers). The sum of the products for all of the organs is

the effective whole-body dose (effective dose).²⁰¹ The estimated risk weighting factors have recently been revised, and a number of additional tissues found in the head and neck region have been included (most importantly the salivary glands, lymphatic nodes, muscle, and oral mucosa).¹⁹⁷ These modifications have resulted in substantial increases (ranging from 32% to 422%) in effective doses for specific maxillofacial radiographic procedures.¹⁷⁷

The effective dose for CBCT radiographic imaging used for orthodontic records is of particular concern, especially as the modal age for initiating orthodontic treatment represents a pediatric population. The radiation risk to ionizing radiation is greater for young children than for adolescents and adults because: 1) the rate of cellular growth and organ development (when radiosensitivity is highest) is greater in young children; 2) children have longer life expectancies, so the cumulative effects of radiation exposures have longer time periods in which they can cause cancers; 3) with CBCT imaging, specific organ and effective doses, (particularly the salivary glands) are, on average, 30% higher for young children than for adolescents¹⁸³; and 4) unless specific, pediatric, exposure–reduction techniques are incorporated, the radiation doses for children (small patients) may exceed typical adult radiation levels (with some currently available CBCT units, it is not possible to implement exposure–reduction techniques). In sum, it is estimated that children may be two to ten times or more prone to radiation-induced carcinogenesis than mature adults.^{175,200-202} Because it is important to consider the increased risks associated with exposing children to ionizing radiation, the American College of Radiology (ACR) has incorporated pediatric, effective-dose estimates in relative radiation level (RRL) designations for specific imaging procedures (Table II).²⁰³ In addition, there are at least two national radiation safety initiatives to raise awareness of using lower radiation doses to image children: Image Gently²⁰⁴ and the National Children’s Dose Registry.²⁰⁵ The AAOMR sought, and received, permission to adopt the ACR, relative-radiation-level designations for several reasons: First, this scheme provides a relative assessment of radiation dose risk based on the premise that with an exposure of 10,000 μ Sv, there is a risk of 1 in 1000 individuals developing cancer; second, the risk is related to diagnostic imaging only (and is unrelated to considerations of background radiation exposure); and three, risk assessment incorporates increased pediatric radiation sensitivity considerations.

For all imaging procedures using ionizing radiation, the clinical benefits should be balanced against the potential radiation risks, which are determined by the relative radiosensitivity of those being imaged and the abilities of the operators to control radiation exposures.

Table II. Estimations of relative radiation level designations for children and adults for orthodontic imaging (with permission from ACR,* 2011)

Relative radiation level	Effective dose estimate range (μ Sv)	
	Adult	Child [†]
0	0	0
☺	<100	<30
☺☺	100-1000	30-300
☺☺☺	1000-10,000	300-3000
☺☺☺☺	10,000-30,000	3,000-10,000

*Some of the information in this document was provided with permission from the American College of Radiology (ACR) and taken from the ACR Appropriateness Criteria. The ACR is not responsible for any deviations from original ACR Appropriateness Criteria content.

[†]Child is defined as any individual less than 18 years of age.

GUIDELINES FOR CBCT IN ORTHODONTICS

The choice of modality used for imaging an orthodontic patient is based on a risk/benefit assessment (i.e., the risk to the patient attributable to radiation exposure in relationship to the benefit to the patient from imaging procedure). Assessment of clinical benefit is primarily patient and practitioner dependent but should be based on the application of sound imaging selection principles. As part of this position statement, the following guidelines are suggested for the use of CBCT in orthodontics:

1. Image appropriately according to clinical condition
2. Assess the radiation dose risk
3. Minimize patient radiation exposure
4. Maintain professional competency in performing and interpreting CBCT studies

1. Image appropriately according to clinical condition

Recently the American Dental Association Council on Scientific Affairs issued an advisory statement on the use of CBCT in dentistry. The AAOMR contributed to the statement,⁶ which is based on the ALARA principle and acknowledges the increased sensitivity of pediatric patients to ionizing radiation and recognizes that patients present with varying degrees of orthodontic complexity. The panel recommends the following general strategies for the use of CBCT in orthodontics:

Recommendation 1.1. The decision to perform a CBCT examination is based on the patient’s history, clinical examination, available radiographic imaging, and the presence of a clinical condition for which the benefits to the diagnosis and/or treatment plan outweigh the potential risks of exposure to radiation, especially in the case of a child or young adult.

Recommendation 1.2. Use CBCT when the clinical question for which imaging is required cannot be answered adequately by lower-dose conventional

dental radiography or alternate non-ionizing imaging modalities.

Recommendation 1.3. Avoid using CBCT on patients to obtain data that can be provided by alternate non-ionizing modalities (e.g., to produce virtual orthodontic study models).

Recommendation 1.4. Use a CBCT protocol that restricts the field of view (FOV), minimizes exposure (mA and kVp), the number of basis images, and resolution yet permits adequate visualization of the region of interest.

Recommendation 1.5. Avoid taking a CBCT scan solely to produce a lateral cephalogram and/or panoramic view if the CBCT would result in higher radiation exposure than would conventional imaging.

Recommendation 1.6. Avoid taking conventional 2D radiographs if the clinical examination indicates that a CBCT study is indicated for proper diagnosis and/or treatment planning or if a recent CBCT study is available.

To assist clinicians in defining the scope of orthodontic conditions and the most appropriate CBCT imaging in each circumstance, specific imaging selection recommendations for the use of CBCT in orthodontics are given in Table III. The proposed recommendations include the phase of treatment (pre-, during-, or post-treatment), the treatment difficulty and the presence of additional skeletal and dental conditions. The table rows list orthodontic phases of treatments and treatment difficulty categories and columns list dental and skeletal clinical conditions. Within each cell, the overall suitability of the CBCT procedure (Table I) and most appropriate FOV are provided. Table IV describes the three FOV ranges most commonly encountered in orthodontic imaging. The concerns in selecting a CBCT FOV are the inclusion of the region of clinical importance and the collimation of the radiation beam to that specific region. The rational for orthodontic image selection recommendations is in Appendix B.

2. Assess the radiation dose risk

Orthodontists must be knowledgeable of the radiation risk of performing CBCT and be able to communicate this risk to their patients. Radiation risk has most often been estimated by calculating the effective dose²⁰¹ of a CBCT scan and comparing this value to; 1) measurements obtained from comparable imaging modalities (e.g., multiples of typical panoramic images or a multi-slice medical CT), 2) background equivalent radiation time (e.g., days of background), or 3) radiation detriment [e.g., probability of x cancers per million scans (stochastic-cancer rate)]. Often the base unit of these comparisons (typical panoramic dose, background radiation, weighted probabilities of fatal and nonfatal cancers) is variable and not absolute. This means, for example, that depending on the panoramic

Table III. Imaging selection recommendations for the use of cone beam computed tomography in orthodontics

Presentation	Dental and skeletal clinical conditions									
	Treatment difficulty	None	Dental structure anomalies	Anomalies in dental position	Compromised dento-alveolar boundaries	Asymmetry	Anteroposterior discrepancies	Vertical discrepancies	Transverse discrepancies	TMJ signs and/or symptoms
Pretreatment	Mild	III	FOV _s (I)	FOV _s (I)	FOV _{s,sm} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{s,sm} (III)
	Moderate	FOV _{m,l} (II)	FOV _s (I)	FOV _{s,sm} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)
	Severe	FOV ₁ (II)	FOV _s (I)	FOV _{s,sm} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)
During treatment		III	FOV _s (III)	FOV _s (II)	FOV _{s,sm} (II)	Presurgical FOV _{m,l} (I)	Presurgical FOV _{m,l} (II)	Presurgical FOV _{m,l} (II)	Presurgical FOV _{m,l} (II)	FOV _{m,l} (II)
	Posttreatment	III	FOV _s (III)	FOV _s (III)	FOV _{s,sm} (III)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)	FOV _{m,l} (II)

CBCT, cone beam computed tomography; Field of View (FOV); FOV_s = Small FOV CBCT imaging; FOV_m = Medium FOV CBCT imaging; FOV₁ = Large FOV CBCT imaging. Consensus Recommendations: I = Likely Indicated; II = Possibly Indicated; III = Likely Not Indicated.

Table IV. Definition of cone beam computed tomography field of view (FOV) ranges for orthodontic imaging

FOV	Abbreviation	Definition
Small	FOV _s	A region of radiation exposure that is limited to a few teeth, a quadrant, and up to two dental arches and that has a spherical volume diameter or cylinder height ≤10 cm.
Medium	FOV _m	A region of radiation exposure that includes the dentition of at least one arch up to both dental arches and that has a spherical volume diameter or cylinder height >10 cm and ≤15 cm.
Large	FOV _l	A region of radiation exposure that includes the TMJ articulations and anatomic landmarks necessary for quantitative cephalometric and/or airway assessment and that has a spherical volume diameter or cylinder height >15 cm.

image dose used for the comparison (e.g., equipment manufacturer and model, film vs. digital acquisition) the risk for CBCT may be reported either conservatively or liberally compared to panoramic radiography.

To standardize comparison of radiation dose risk between various imaging procedures, this position statement recommends the use of RRLs (Table II). The RRL for various imaging examinations used either as an isolated procedure or for a course of orthodontics can be determined for adults and children using published effective dose calculations (Table VI).^{90,159,176-196,206,207} Calculations of RRL levels in millisieverts (mSv; 1mSv = 1000 μSv) were made with methods described elsewhere,¹⁹⁷ and data from the 7th Biological Effects of Ionizing Radiation report.²⁰⁸ The estimate in the report, and the basis for subsequent levels of radiation risk, is that approximately 1 in 1000 individuals develop cancer from an exposure of 10,000 μSv.¹⁹⁷ RRL assignments are based on reviews of current literature. These assignments are revised periodically, as practice evolves and further information becomes available.

Based on these considerations, the following recommendations are suggested for assessing patient radiation dose risk for CBCT in orthodontics:

Recommendation 2.1. Consider the RRL (Table II) when assessing the imaging risk for imaging procedures over a course of orthodontic treatment. Table V contains

the effective doses for specific orthodontic protocols and various modalities. Appendix C provides an example of the calculation of RRL for Orthodontic Imaging.

Recommendation 2.2. Because CBCT exposes patients to ionizing radiation that may pose elevated risks to some patients (pregnant or younger patients), explain and disclosure to patients radiation exposure risks, benefits and imaging modality alternatives and document this in the patients' records.

3. Minimize patient radiation exposure

Depending on the equipment type and operator preferences, operators can alter radiation doses to patients by adjusting various exposure (e.g., milliamperage, kilovoltage), image-quality (e.g., number of basis images, resolution, arc of trajectory) and beam-collimation (e.g., FOV) settings. CBCT units from different manufacturers vary in dose by as much as 10-fold for an equivalent FOV examination (Table V).¹⁸⁴ In addition, adjustments of exposure factors to improve image quality are available in many CBCT units and can cause as much as 7-fold differences in patient doses (Table V).¹⁸⁴ If CBCT imaging is warranted, appropriate selection of the FOV to match the region of interest (ROI) may provide a substantial dose savings.

Based on these considerations, the following specific recommendations are made to minimize patient radiation exposure for CBCT in orthodontics:

Recommendation 3.1. Perform CBCT imaging with acquisition parameters adjusted to the nominal settings consistent with providing appropriate images of task-specific diagnostic quality for the desired diagnostic information required: 1) Use a pulsed exposure mode of acquisition, 2) Optimize exposure settings (mA, kVp), 3) Reduce the number of basis projection images, and 4) Employ dose reduction protocols (e.g., reduced resolution) when possible.

Recommendation 3.2. When other factors remain the same, reduce the size of the FOV to match the ROI; however, selection of FOV may result in automatic or default changes in other technical factors (e.g., mAs) that should be considered because these concomitant changes can result in an increase in dose.

Recommendation 3.3. Use patient protective shielding (such as, lead torso aprons and consider the use of thyroid shields) when possible (e.g., maxillary only scan), to minimize exposure to radiosensitive organs outside the FOV of the exposure.

Recommendation 3.4. Ensure that all CBCT equipment is properly installed, routinely calibrated and updated, and meets all governmental requirements and regulations.

Appendix C provides an example of the calculation of the RRL for both adults and children with and

Table V. Selected published effective doses ($E_{ICRP, 2007}$) in microSieverts [μSv] for various field of view (FOV) cone beam computed tomography devices used in orthodontics in comparison with multi-slice computed tomography (MSCT), rotational panoramic and cephalometric radiography

Examination	CBCT unit	Scanning volume (cm ²)	Protocol	E (μSv) ^{Reference}
Large FOV CBCT (>15 cm height/diameter)	3DeXAM	17 × 23	0.4 mm resolution	72 ¹⁹⁶
	3D Accuitomo 170	17 × 12	Adolescent; 10 years old	216 ¹⁸³ ; 282 ¹⁸³
	CB MercuRay	15 × 15	Maxillofacial/TMJ	436 ¹⁸⁴ ; 569 ¹⁸⁴ ; 680 ¹⁹⁵ ; 511 ¹⁸⁰ /436 ⁹⁰
		20 × 20	SR/HR/TMJ	558 ¹⁷⁷ ; 761 ¹⁹⁵ /1025 ¹⁷⁷ ; 1073 ¹⁸⁴ /916 ⁹⁰
	Galileos	15 × 15	High/low dose	128 ¹⁸⁴ /70 ¹⁸⁴
	Galileos Comfort	15 × 15	Adult; adolescent; 10 years old	84 ¹⁹¹ ; 71 ¹⁸³ ; 70 ¹⁸³
	i-CAT Classic	16 × 22	Low/high resolution	65-69 ¹⁹² ; 193 ¹⁷⁷ ; 82 ¹⁷⁸ ; 206 ¹⁸⁶ ; 110 ¹⁸¹ /127-131 ¹⁹²
	i-CAT Next Generation	23 × 17	Standard/ultra	74 ¹⁸⁴ ; 78 ¹⁹⁰
	Iluma	19 × 19		98 ¹⁸⁴ /498 ¹⁸⁴
	Iluma Elite	21 × 14	With; without filtration	368 ¹⁹¹
	KODAK 9500	18 × 20		136 ¹⁹¹ ; 166 ¹⁸⁸ /260 ¹⁸⁸
	NewTom 3G	15 × 15/20 × 20	Adult; adolescent; 10 years old	57 ¹⁷⁸ /59 ¹⁷⁷ ; 68 ¹⁸⁴
	NewTom 9000	15 × 15		56 ¹⁵⁹ ; 95 ¹⁹³ ; 52 ¹⁸⁴
	Newtom VGi	15 × 15		194 ¹⁹¹
	Skyview 3D	17 × 17		87 ¹⁹¹ ; 90 ¹⁸³ ; 105 ¹⁸³
Medium FOV CBCT (>10 cm and ≤15 cm height/diameter)	3DeXAM	13 × 16		0.3 mm resolution
	3D Accuitomo 170	10 × 14	Adolescent; 10 years old	188 ¹⁸³ ; 237 ¹⁸³
	CB Mercuray	10 × 10	Maxillofacial/TMJ imaging	283 ¹⁷⁷ ; 407 ¹⁸⁴ ; 603 ¹⁹⁵ /283 ⁹⁰
	i-CAT Classic	13 × 16	Adult; adolescent; 10 years old	61 ¹⁵⁹ ; 105 ¹⁷⁷ ; 134 ¹⁸⁶ ; 69 ¹⁸⁴
	i-CAT Next Generation	13 × 16		87 ¹⁸⁴ ; 83 ¹⁹¹ ; 77 ¹⁹⁰ ; 82 ¹⁸³ ; 134 ¹⁸³
	NewTom VG	11 × 15	Adult; adolescent; 10 years old	83 ¹⁹¹ ; 81 ¹⁸³ ; 114 ¹⁸³
	Scanora 3D	13.5 × 14.5	Adult; adolescent; 10 years old	68 ¹⁹¹ ; 74 ¹⁸³ ; 85 ¹⁸³
	Small FOV CBCT (≤10 cm height/diameter)	3DeXAM	5 × 10	Man
		8 × 16	0.25; 0.30 resolution	170 ¹⁹⁶ ; 45 ¹⁹⁶
		4 × 16	Max 0.125 mm; 0.3 mm resolution/man 0.125 mm; 0.3 mm resolution	68 ¹⁹⁶ ; 33 ¹⁹⁶ /76 ¹⁹⁶ ; 38 ¹⁹⁶
		8 × 8	0.125 mm; 0.3 mm resolution	122 ¹⁹⁶ ; 62 ¹⁹⁶
3D Accuitomo IID		3 × 4		27 ¹⁷⁹
3D Accuitomo FPD		4 × 4/6 × 6		102 ¹⁸⁰ ; 20 ¹⁸⁵ /43 ¹⁸⁵ ; 50 ¹⁸⁰ ; 166 ¹⁷⁹

Table V. Continued

<i>Examination</i>	<i>CBCT unit</i>	<i>Scanning volume (cm²)</i>	<i>Protocol</i>	<i>E (μSv)^{Reference}</i>
3D Accuitomo 170		4 × 4	Man adult; adolescent; 10 year old	43 ¹⁹¹ ; 32 ¹⁸³ ; 28 ¹⁸³
		5 × 10	Max	54 ¹⁹¹
		5 × 14	Max adolescent; 10 years old	70 ¹⁸³ ; 214 ¹⁸³
AZ3000CT		7.9 × 7.1		333 ¹⁸²
i-CAT Classic		6 × 16	Man SR; HR/Mx SR; HR	96 ¹⁸⁶ ; 189 ¹⁸⁶ /59 ¹⁸⁶ ; 93 ¹⁸⁶
i-CAT Next Generation		6 × 16	Man SR; HR/Max SR; HR	74 ¹⁸⁴ ; 45 ¹⁹¹ ; 58 ¹⁹⁰ ; 113 ¹⁹⁰ / 32 ¹⁹⁰ ; 60 ¹⁹⁰
Implagraphy KODAK 9500		6 × 16	Max adolescent; 10 year old/ man adolescent; 10 year old	33 ¹⁸³ ; 43 ¹⁸³ /49 ¹⁸³ ; 63 ¹⁸³
		5 × 8		83 ¹⁸²
KODAK 9000 3D		5 × 15/9 × 15	Without; with filtration	93 ¹⁸⁸ ; 76 ¹⁸⁸ /92 ¹⁹¹ ; 163 ¹⁸⁸ ; 98 ¹⁸⁸
Newtom VGi Pan eXam Plus 3D		5 × 3.7	Max anterior adult; 10 years old/man molar adult; adolescent	19 ¹⁹¹ ; 16 ¹⁸³ /40 ¹⁹¹ ; 24 ¹⁸³
		8 × 12		265 ¹⁹¹
Pan eXam Plus 3D		4.1 × 6.1	Max 0.133 mm; 0.2 mm resolution/man 0.133 mm; 0.2 mm resolution	79 ¹⁹⁶ ; 40 ¹⁹⁶ /115 ¹⁹⁶ ; 49 ¹⁹⁶
		7.8 × 6.4	Max 0.2 mm; 0.3 mm resolution/man 0.2 mm; 0.3 mm resolution	125 ¹⁹⁶ ; 79 ¹⁹⁶ /184 ¹⁹⁶ ; 110 ¹⁹⁶
Picasso Trio PreXion		7 × 12	Low/high dose	81 ¹⁹¹ /123 ¹⁹¹
		8.1 × 7.6	High/standard resolution	388 ¹⁸⁴ /189 ¹⁸⁴
		6 × 16	Max adolescent; 10 years old/ man adolescent, 10 years old	33 ¹⁸³ ; 43 ¹⁸³ /49 ¹⁸³ ; 63 ¹⁸³
ProMax 3D		8 × 8	High/standard/low	674 ¹⁷⁹ ; 652 ¹⁸⁴ ; 122 ¹⁹¹ ; 306 ¹⁹³ /197 ¹⁹³ /488 ¹⁸⁴ ; 30 ¹⁸⁷ ; 28 ¹⁹¹
		8 × 8	Adolescent; 10 years old	18 ¹⁸³ ; 28 ¹⁸³
Pax-Uni3D		5 × 5	Max anterior	44 ¹⁹¹
Scanora 3D		6 × 6		91 ¹⁷⁹
		7.5 × 10	Max/man/both	46 ¹⁹¹ /47 ¹⁹¹ /45 ¹⁹¹
Veraviewepocs 3D		7.5 × 10	Adolescent; 10 year old	52 ¹⁸³ ; 67 ¹⁸³
		4 × 4/4 × 8/6 × 6/8 × 8		31 ¹⁸⁵ /40 ¹⁸⁵ /40 ¹⁸⁵ /73 ¹⁹¹

(continued on next page)

Table V. Continued

Examination	CBCT unit	Scanning volume (cm ²)	Protocol	E (μSv) ^{Reference}	
MSCT	Siemens Somatom	Lower jaw/head	Head sensation 16; volume zoom 4	474 ¹⁷⁸ ; 494 ¹⁷⁸ /995 ¹⁷⁸ ; 1110 ¹⁷⁸	
		Lower jaw	Sensation 10; emotion 6	426 ¹⁸² ; 199 ¹⁸²	
		10 × 12	Sensation 64	430 ¹⁵⁹ ; 860-534 ¹⁷⁷	
		20 × 12.8/11.7	Sensation 64 adolescent; 10 years old	1047 ¹⁸³ ; 605 ¹⁸³	
	Philips Mx8000IDT	34.8 × 25	25 × 41.25	Lower jaw; head	541 ¹⁷⁸ ; 1160 ¹⁷⁸
				GE 4 Slice CT	685 ¹⁷⁹
				GE 64 Slice CT	1410 ¹⁷⁹
				Toshiba Aquilion 64	990 ¹⁸¹
				HiSpeed QX/I	769 ¹⁸⁰
				Panoramic	Planmeca Promax
Planmeca PM Proline 2000	38 ²⁰⁷ ; 12 ²⁰⁷				
Veraviewepocs	6 ¹⁸³				
Sirona Orthophos	10 ¹⁵⁹ ; 50 ¹⁸¹				
Cephalometric	Instrumentarium OP100	30 × 15	Lat ceph	21.5 ¹⁹²	
				PSP	5.6 ¹⁸⁴
	Orthophos DS	18 × 15	Lat ceph	10 ¹⁵⁹	
				Instrumentarium OC 100	4.5 ¹⁹²
	Veraviewepocs 2D	20 × 20	Lat ceph	2 ¹⁸³	
				Planmeca Promax PA	5.1 ¹⁸⁴

CBCT, cone beam computed tomography; PSP, photo-stimulable phosphor; CCD, charged coupled device-based technology; Max, maxillary; Man, mandibular; TMJ, temporomandibular joint; MSCT, multi-slice computed tomography; HR, high resolution; SR, standard resolution; Lat ceph, lateral cephalometric image; PA, posteroanterior cephalometric image; N/A, not available.

Product/Manufacturer details: 3DeXAM (KaVo Dental GmbH, Biberach/Riß, Germany); 3D Accuitomo 170 (J. Morita Mfg. Corp., Kyoto, Japan); CB Mercuray (Hitachi Medical Systems, Kyoto, Japan); Galileos (Sirona Dental Systems GmbH, Bensheim Germany); Galileos Comfort (Sirona Dental Systems GmbH, Bensheim Germany); i-CAT Classic (Imaging Sciences International, Hatfield, PA); i-CAT Next Generation (Imaging Sciences International, Hatfield, PA); Iluma (Imtec (3M), Ardmore, OK); Iluma Elite (Imtec (3M), Ardmore, OK); KODAK 9500 (Kodak Dental Systems, Carestream Health, Rochester, NY); NewTom 3G (Quantitative Radiology, Verona, Italy); NewTom 9000 (Quantitative Radiology, Verona, Italy); NewTom VGi (Quantitative Radiology, Verona, Italy); Skyview 3D (MyRay, Cefla Dental Group, Imola, Italy); 3DeXAM (KaVo Dental GmbH, Biberach/Riß, Germany); 3D Accuitomo 170 (J. Morita Mfg. Corp., Kyoto, Japan); CB Mercuray (Hitachi Medical Systems, Kyoto, Japan); i-CAT Classic (Imaging Sciences International, Hatfield, PA); i-CAT Next Generation (Imaging Sciences International, Hatfield, PA); NewTom VG (Quantitative Radiology, Verona, Italy); Scanora 3D (Soredex, Tuusula, Finland); 3DeXAM (KaVo Dental GmbH, Biberach/Riß, Germany); 3D Accuitomo IID (J. Morita Mfg. Corp., Kyoto, Japan); 3D Accuitomo FPD (J. Morita Mfg. Corp., Kyoto, Japan); 3D Accuitomo 170 (J. Morita Mfg. Corp., Kyoto, Japan); AZ3000CT (Asahi Roentgen, Kyoto, Japan); i-CAT Classic (Imaging Sciences International, Hatfield, PA); i-CAT Next Generation (Imaging Sciences International, Hatfield, PA); Implagraphy (Vatech, E-WOO Technology Co, Ltd. Republic of Korea); KODAK 9500 (Kodak Dental Systems, Carestream Health, Rochester, NY); KODAK 9000 3D (Kodak Dental Systems, Carestream Health, Rochester, NY); NewTom VGi (Quantitative Radiology, Verona, Italy); Pan eXam Plus 3D (Palodex Group Oy, Tuusula, Finland); Picasso Trio (Vatech, Co, Ltd. Republic of Korea); PreXion 3D (PreXion Inc., San Mateo, CA); ProMax 3D (Planmeca OY, Helsinki, Finland); Pax-Uni3D (Vatech, Technology Co, Ltd. Republic of Korea); Scanora 3D (Soredex, Tuusula, Finland); Veraview epocs 3D (J. Morita Mfg. Corp., Kyoto, Japan); Siemens Somatom (Siemens Medical Solutions USA, Malvern, PA); Philips Mx8000IDT (Philips Medical Systems, Best, the Netherlands); GE 4 slice CT (GE Medical Systems, Little Chalfont, UK); GE 64 slice CT (GE Medical Systems, Little Chalfont, UK); Toshiba Aquilion 64 (Toshiba Medical Systems Corporation, Tochigi, Japan); HiSpeed QX/I (GE Medical Systems, Little Chalfont, UK); Planmeca Promax (Planmeca, Helsinki, Finland); Planmeca PM Proline 2000 (Planmeca, Helsinki, Finland); Veraview epocs (J. Morita Mfg. Corp., Kyoto, Japan); Sirona Orthophos (Sirona Dental Systems GmbH, Bensheim Germany); Instrumentarium OP100 (Instrumentarium Dental, Tuusula, Finland); Orthophos DS (Sirona Dental Systems GmbH, Bensheim Germany); Instrumentarium OC 100 (Instrumentarium Dental, Tuusula, Finland); Veraview epocs 2D (J. Morita Mfg. Corp., Kyoto, Japan); Planmeca Promax PA (Planmeca OY, Helsinki, Finland).

Table VI. Examples of the calculation of the RRL associated with specific imaging protocols used in orthodontics

Protocol	Modality	Stage of treatment			Dose (μSv)		Relative radiation level*	
		Initial diagnostic	Mid-treatment	Post-treatment	Sub-total	Total	Child	Adult
Conventional imaging	Panoramic [†]	+	+	+	36	47.2	⊗⊗	⊗
	Lateral cephalogram [‡]	+	–	+	11.2			
Conventional + small FOV CBCT	Panoramic [†]	+	+	+	36	107.2	⊗⊗	⊗
	Lateral cephalogram [‡]	+	–	+	11.2			
	Small FOV CBCT [§]	+	–	–	60			
Large FOV CBCT + conventional imaging	Panoramic [†]	–	+	+	24	112.6	⊗⊗	⊗⊗
	Lateral cephalogram [‡]	–	–	+	5.6			
	Large FOV CBCT	+	–	–	83			
Large FOV CBCT	Large FOV CBCT	+	+	+	249	249	⊗⊗	⊗⊗

CBCT, cone beam computed tomography; FOV, field of view; CCD, charged coupled device technology; Sub-total, product of the times when the modality is used at each stage over a course of treatment by the average effective dose per modality exposure; Total, sum of subtotals for a particular orthodontic imaging protocol.

*American College of Radiology relative radiation level²⁰³; ⊗, child (<30 μSv), adult (<100 μSv); ⊗⊗, child (<30-300 μSv), adult (100-1000 μSv).

[†]Planmeca PM Proline 2000 (low dose) – charged coupled device (12 μSv).²⁰⁷

[‡]Photostimulable storage phosphor (5.6 μSv).¹⁷⁷

[§]i-CAT Next Generation – Maxilla 6 cm FOV height, high resolution (60 μSv).¹⁹⁰

^{||}i-CAT Next Generation – 16 × 13 cm (83 μSv).¹⁹¹

without CBCT imaging for representative orthodontic imaging protocols (Table VI).

4. Maintain professional competency in performing and interpreting CBCT studies

Orthodontists must be able to exercise judgment by applying professional standards to all aspects of CBCT. Any radiographic image prescribed and/or performed by a dental practitioner may contain information that is important to the management or general health of the patient. Incidental findings in CBCT images of orthodontic patients are common,²⁰⁹⁻²¹³ and some are critical to patient health.²¹⁴ Clinicians who order or perform CBCT for orthodontic patients are responsible for interpreting the entire image volumes, just as they are responsible for interpreting all regions of other radiographic images that they order.^{215,216}

Based on these considerations, the following recommendations are related to performing and interpreting CBCT studies:

Recommendation 4.1. Clinicians have an obligation to attain and improve their professional skills through lifelong learning in regards to performing CBCT examinations as well as interpreting the resultant images. Clinicians need to attend continuing education courses (such as those offered by the American Dental Association Continuing Education Recognition Program) to maintain familiarity with the technical and operational aspects of CBCT and to maintain current knowledge of scientific advances and health risks associated with the use of CBCT.

Recommendation 4.2. Clinicians have legal responsibilities when operating CBCT equipment and interpreting images and are expected to comply with all governmental and third party payer (e.g., Medicare) regulations.

Recommendation 4.3. It is important that patients/guardians know about the limitations of CBCT with regard to visualization of soft tissues, artifacts and noise.

EMERGING DEVELOPMENTS

CBCT acquisition technology continues to develop and a number of innovations are proposed to improve image quality, increase utility and reduce radiation output. These include the use of automatic exposure control with photon counting, added filtration, flat panel detectors with greater photon sensitivity, customizable FOV collimation, variable exposure parameters (mA, kVp) and image quality settings (e.g., scan trajectory options and number of basis images). The image quality and dose reductions purported by such innovations should be assessed critically and verified by independent published research.

SUMMARY

The recommendations provided for the use of CBCT in orthodontics are neither rigid guidelines nor do they represent or imply a standard of care. While it is the responsibility of each practitioner to make a decision, along with the patient/family, as to what imaging is considered to be in the patient’s best interest, this

position statement is intended to assist the clinician in the decision making process.

This position statement supports and affirms the position of the American Dental Association Council on Scientific Affairs in that the selection of CBCT imaging should be based on initial clinical evaluation and must be justified based on individual need.⁶ The perceived or actual benefits to the patient must outweigh the radiation risks. Exposure of patients to ionizing radiation must never be considered "routine." It is important to perform a thorough clinical examination prior to performing or ordering any radiographic study. This position statement provides four guidelines for CBCT use in orthodontic practice: 1) Image appropriately by applying imaging selection recommendations, 2) Assess the radiation dose risk, 3) Minimize patient radiation exposure and, 4) Maintain professional competency in performing and interpreting CBCT studies.

Some of the information in this document was provided with permission from the ACR and taken from the ACR Appropriateness Criteria. The ACR is not responsible for any deviations from original ACR Appropriateness Criteria content. The panel gratefully acknowledges the contributions of Dr. Michael M. Bornstein, Department of Oral Surgery and Stomatology, School of Dental Medicine, University of Bern, Bern, Switzerland and Professor Reinhilde Jacobs, Oral Imaging Center, Department of Oral Health Sciences, KU Leuven & Dentistry, University Hospitals Leuven, Belgium for assistance in the development of Table V.

Panel members:

Carla A. Evans (Co-Chair)
William C. Scarfe (Co-Chair)
Mansur Ahmad
Lucia H.S. Cevidanes
John B. Ludlow
J. Martin Palomo
Kirt E. Simmons
Stuart C. White

REFERENCES

- Baumrind S, Miller D, Molthen R. The reliability of head film measurements. 3. Tracing superimposition. *Am J Orthod.* 1976;70:617-644.
- Moyers RE, Bookstein FL. The inappropriateness of conventional cephalometrics. *Am J Orthod.* 1979;75:599-617.
- Johnston LE Jr. A few comments on an elegant answer in search of useful questions. *Semin Orthod.* 2011;17:13-14.
- Little RM, Wallen TR, Riedel RA. Stability and relapse of mandibular anterior alignment-first premolar extraction cases treated by traditional edgewise orthodontics. *Am J Orthod.* 1981;80:349-365.
- National Council on Radiation Protection & Measurements. *Radiation Protection in Dentistry (Report No. 145)*. Bethesda, MD: NCRP Publications; 2003.
- American Dental Association Council on Scientific Affairs. The use of cone-beam tomography in dentistry. An advisory statement from the American Dental Association Council on Scientific Affairs. *J Am Dent Assoc.* 2012;143:899-902.
- Matteson SR, Joseph LP, Bottomley W, et al. The selection of patients for X-ray examinations: dental radiographic examinations. In: *Center for Devices and Radiological Health*, ed. U.S. Department of Health and Human Services, Public Health Service, Food and Drug Administration; 1987.
- U.S. Department of Health and Human Services, Public Health Service, Food and Drug Administration; and American Dental Association, Council on Dental Benefit Programs, Council on Scientific Affairs. *The Selection of Patients for Dental Radiographic Examinations*. Rev. ed. 2004. Available at: www.ada.org/prof/resources/topics/radiography.asp. Accessed May 26, 2012.
- American Dental Association Council on Scientific Affairs. The use of dental radiographs: update and recommendations. *J Am Dent Assoc.* 2006;137:1304-1312.
- Janssens A, Horner K, Rushton V, et al. *Radiation Protection: European Guidelines on Radiation Protection in Dental Radiology—the Safe Use of Radiographs in Dental Practice*, 2003. Available at: www.sefm.es/docs/otros/raddigUE.pdf. Accessed April 20, 2012.
- SEDENTEXCT Project. Chapter 4, Justification and referral criteria. The developing dentition. In: *Radiation Protection: Cone Beam CT for Dental and Maxillofacial Radiology. Evidence Based Guidelines 2011(v2.0 Final)*. 2011:36-48. Available at: http://www.eadmfr.info/sites/default/files/guidelines_final.pdf. Accessed January 14, 2013.
- European Commission. *Item 4.2 the Developing Dentition in Protection Radiation No. 172. Cone Beam CT for Dental and Maxillofacial Radiology (Evidence-based Guidelines)*. 2011:45-56. Available at: http://ec.europa.eu/energy/nuclear/radiation_protection/doc/publication/172.pdf. Accessed January 14, 2013.
- Isaacson KG, Thom AR, Horner K, Whaites E. *Orthodontic Radiographs—Guidelines for the Use of Radiographs in Clinical Orthodontics*. 3rd ed. London: British Orthodontic Society; 2008.
- Müssig E, Wörtche R, Lux CJ. Indications for digital volume tomography in orthodontics. *J Orofac Orthop.* 2005;66:241-249.
- Hechler SL. Cone-beam CT: applications in orthodontics. *Dent Clin N Am.* 2008;52:753-759.
- White SC, Pae EK. Patient image selection criteria for cone beam computed tomography imaging. *Semin Orthod.* 2009;15:19-28.
- Merrett SJ, Drage NA, Durning P. Cone beam computed tomography: a useful tool in orthodontic diagnosis and treatment planning. *J Orthod.* 2009;36:202-210.
- Mah JK, Huang JC, Choo H. Practical applications of cone-beam computed tomography in orthodontics. *J Am Dent Assoc.* 2010;141(suppl 3):7S-13S.
- Kapila S, Conley RS, Harrell WE Jr. The current status of cone beam computed tomography imaging in orthodontics. *Dentomaxillofac Radiol.* 2011;40:24-34.
- Nervina JM. Cone beam computed tomography use in orthodontics. *Aust Dent J.* 2012;57(suppl 1):95-102.
- Peck JL, Sameshima GT, Miller A, Worth P, Hatcher DC. Mesiodistal root angulation using panoramic and cone beam CT. *Angle Orthod.* 2007;77:206-213.
- Liu DG, Zhang WL, Zhang ZY, Wu YT, Ma XC. Three-dimensional evaluations of supernumerary teeth using cone-beam computed tomography for 487 cases. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2007;103:403-411.

23. Treil J, Braga J, Loubes JM, et al. 3D tooth modeling for orthodontic assessment. *Semin Orthod.* 2009;15:42-47.
24. Dudic A, Giannopoulou C, Leuzinger M, Kiliaridis S. Detection of apical root resorption after orthodontic treatment by using panoramic radiography and cone-beam computed tomography of super-high resolution. *Am J Orthod Dentofacial Orthop.* 2009;135:434-437.
25. Liedke GS, Dias de Silveira HE, Dias de Silveira HL, Dutra V, Poli de Figueiredo JA. Influence of voxel size in the diagnostic ability of cone beam tomography to evaluate simulated external root resorption. *J Endod.* 2009;35:233-235.
26. Sherrard JF, Rossouw PE, Benson BW, Carrillo R, Buschang PH. Accuracy and reliability of tooth and root lengths measured on cone-beam computed tomographs. *Am J Orthod Dentofac Orthop.* 2010;137(4 suppl):S100-S108.
27. Katheria BC, Kau CH, Tate R, Chen JW, English J, Bouquot J. Effectiveness of impacted and supernumerary tooth diagnosis from traditional radiography versus cone beam computed tomography. *Ped Dent.* 2010;32:304-309.
28. Leuzinger M, Dudic A, Giannopoulou C, Killaridis S. Root-contact evaluation by panoramic radiography and cone-beam computed tomography of super-high resolution. *Am J Orthod Dentofac Orthop.* 2010;137:389-392.
29. Lund H, Grondahl K, Grondahl HG. Cone beam computed tomography for assessment of root length and marginal bone level during orthodontic treatment. *Angle Orthod.* 2010;80:466-473.
30. Van Elslande D, Heo G, Flores-Mir C, Carey J, Major PW. Accuracy of mesiodistal root angulation projected by cone-beam computed tomographic panoramic-like images. *Am J Orthod Dentofac Orthop.* 2010;137(4 suppl):S94-S99.
31. Shemesh H, Cristescu RC, Wesslink PR, Wu MK. The use of cone-beam computed tomography and digital periapical radiographs to diagnose root perforations. *J Endod.* 2011;37:513-516.
32. Makedonas D, Lund H, Grondahl K, Hansen K. Root resorption diagnosed with cone beam computed tomography after 6 months of orthodontic treatment with fixed appliance and the relation to risk factors. *Angle Orthod.* 2012;82:196-201.
33. Chaushu S, Chaushu G, Becker A. The role of digital volume tomography in the imaging of impacted teeth. *World J Orthod.* 2004;5:120-132.
34. Nakajima A, Sameshima GT, Arai Y, Homme Y, Shimizu N, Dougherty H. Two- and three-dimensional orthodontic imaging using limited cone beam-computed tomography. *Angle Orthod.* 2005;75:895-903.
35. Walker L, Enciso R, Mah J. Three-dimensional localization of maxillary canines with cone-beam computed tomography. *Am J Orthod Dentofacial Orthop.* 2005;128:418-423.
36. Kau CH, Richmond S, Palomo JM, Hans MG. Three-dimensional cone beam computerized tomography in orthodontics. *J Orthod.* 2005;32:282-293.
37. Bjerklín K, Ericson S. How a computerized tomography examination changed the treatment plans of 80 children with retained and ectopically positioned maxillary canines. *Angle Orthod.* 2006;76:43-51.
38. Maverna R, Gracco A. Different diagnostic tools for the localization of impacted maxillary canines: clinical considerations. *Prog Orthod.* 2007;8:28-44.
39. Liu D, Zhang W, Zhang Z, Wu Y, Ma X. Localization of impacted maxillary canines and observation of adjacent incisor resorption with cone-beam computed tomography. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2008;105:91-98.
40. Bedoya MM, Park JH. A review of the diagnosis and management of impacted maxillary canines. *J Am Dent Assoc.* 2009;140:1485-1493.
41. Gracco A, Lombardo L, Mancuso G, Gravina V, Siciliani G. Upper incisor position and bony support in untreated patients as seen on CBCT. *Angle Orthodontist.* 2009;79:692-702.
42. Kau CH, Pan P, Gallerano RL, English JD. A novel 3D classification system for canine impactions – the KPG index. *Int J Med Robot.* 2009;5:291-296.
43. Haney E, Gansky SA, Lee JS, et al. Comparative analysis of traditional radiographs and cone-beam computed tomography volumetric images in the diagnosis and treatment planning of maxillary impacted canines. *Am J Orthod Dentofacial Orthop.* 2010;137:590-597.
44. Tamimi D, ElSaid K. Cone beam computed tomography in the assessment of dental impactions. *Semin Orthod.* 2009;15:57-62.
45. Becker A, Chaushu C, Casap-Caspi N. Cone-beam computed tomography and the orthosurgical management of impacted teeth. *J Am Dent Assoc.* 2010;141:14S-18S.
46. Botticelli S, Verna C, Cattaneo PM, Heidmann J, Melsen B. Two- versus three-dimensional imaging in subjects with unerupted maxillary canines. *Eur J Orthod.* 2011;33:344-349.
47. Oberoi S, Knueppel S. Three-dimensional assessment of impacted canines and root resorption using cone beam computed tomography. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2012;113:260-267.
48. Hofmann E, Medelnic J, Fink M, Lell M, Hirschfelder U. Three-dimensional volume tomographic study of the imaging accuracy of impacted teeth: MSCT and CBCT comparison – an in vitro study. *Eur J Orthod.* 2013;35:286-294.
49. Guerrero ME, Shahbazian M, Elsiene Bekkering G, Nackaerts O, Jacobs R, Horner K. The diagnostic efficacy of cone beam CT for impacted teeth and associated features: a systematic review. *J Oral Rehabil.* 2011;38:208-216.
50. Nguyen E, Boychuk D, Orellana M. Accuracy of cone-beam computed tomography in predicting the diameter of unerupted teeth. *Am J Orthod Dentofacial Orthop.* 2011;140:e59-e66.
51. Rungcharassaeng K, Caruso JM, Kan JY, Kim J, Taylor G. Factors affecting buccal bone changes of maxillary posterior teeth after rapid maxillary expansion. *Am J Orthod Dentofacial Orthop.* 2007;132:428.e1-428.e8.
52. Loubele M, Van Assche N, Carpentier K, et al. Comparative localized linear accuracy of small-field cone-beam CT and multislice CT for alveolar bone measurements. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2008;105:512-518.
53. Leung CC, Palomo L, Griffith R, Hans MG. Accuracy and reliability of cone-beam computed tomography for measuring alveolar bone height and detecting bony dehiscences and fenestrations. *Am J Orthod Dentofac Orthop.* 2010;137(4 suppl):S109-S119.
54. Molen AD. Considerations in the use of cone-beam computed tomography for buccal bone measurements. *Am J Orthod Dentofac Orthop.* 2010;137(4 suppl):S130-S135.
55. Timock AM, Cook V, McDonald T, et al. Accuracy and reliability of buccal bone height and thickness measurements from cone-beam computed tomography imaging. *Am J Orthod Dentofacial Orthop.* 2011;140:734-744.
56. Yagci A, Veli I, Uysal T, Ucar FI, Ozer T, Enhos S. Dehiscence and fenestration in skeletal Class I, II, and III malocclusions assessed with cone-beam computed tomography. *Angle Orthod.* 2012;82:67-74.
57. Sievers MM, Larson BE, Gaillard PR, Wey A. Asymmetry assessment using cone beam CT. A Class I and Class II patient comparison. *Angle Orthod.* 2012;82:410-417.

58. AlHadidi A, Cevidanes LH, Mol A, Ludlow J, Styner M. Comparison of two methods for quantitative assessment of mandibular asymmetry using cone beam computed tomography image volumes. *Dentomaxillofac Radiol.* 2011;40:351-357.
59. de Moraes ME, Hollender LG, Chen CS, Moraes LC, Balducci I. Evaluating craniofacial asymmetry with digital cephalometric images and cone-beam computed tomography. *Am J Orthod Dentofacial Orthop.* 2011;139:e523-e531.
60. Damstra J, Fourie Z, Ren Y. Evaluation and comparison of postero-anterior cephalograms and cone-beam computed tomography images for the detection of mandibular asymmetry. *Eur J Orthod.* 2013;35:45-50.
61. Veli I, Uysal T, Ozer T, Ucar FI, Eruz M. Mandibular asymmetry in unilateral and bilateral posterior crossbite patients using cone-beam computed tomography. *Angle Orthod.* 2011;81:966-974.
62. Kook YA, Kim Y. Evaluation of facial asymmetry with three dimensional cone-beam computed tomography. *J Clin Orthod.* 2011;45:112-115.
63. Cevidanes LH, Alhadidi A, Paniagua B, et al. Three-dimensional quantification of mandibular asymmetry through cone-beam computerized tomography. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2011;111:757-770.
64. Park JU, Kook YA, Kim Y. Assessment of asymmetry in a normal occlusion sample and asymmetric patients with three-dimensional cone beam computed tomography: a study for a transverse reference plane. *Angle Orthod.* 2012;82:860-867.
65. Orentlicher G, Goldsmith D, Horowitz A. Applications of 3-dimensional virtual computerized tomography technology in oral and maxillofacial surgery: current therapy. *J Oral Maxillofac Surg.* 2010;68:1933-1959.
66. Tucker S, Cevidanes LH, Styner M, et al. Comparison of actual surgical outcomes and 3-dimensional surgical simulations. *J Oral Maxillofac Surg.* 2010;68:2412-2421.
67. Lagravère MO, Carey J, Heo G, Toogood RW, Major PW. Transverse, vertical, and anteroposterior changes from bone-anchored maxillary expansion vs traditional rapid maxillary expansion: a randomized clinical trial. *Am J Orthod Dentofacial Orthop.* 2010;137:304.e1-304.e12.
68. Cevidanes LH, Tucker S, Styner M, et al. Three-dimensional surgical simulation. *Am J Orthod Dentofacial Orthop.* 2010;138:361-371.
69. Heymann GC, Cevidanes L, Cornelis M, De Clerck HJ, Tulloch JF. Three-dimensional analysis of maxillary protraction with intermaxillary elastics to miniplates. *Am J Orthod Dentofacial Orthop.* 2010;137:274-284.
70. Almeida RC, Cevidanes LH, Carvalho FA, et al. Soft tissue response to mandibular advancement using 3D CBCT scanning. *Int J Oral Maxillofac Surg.* 2011;40:353-359.
71. Gateno J, Xia JJ, Teichgraber JF. New 3-dimensional cephalometric analysis for orthognathic surgery. *J Oral Maxillofac Surg.* 2011;69:606-622.
72. Kim YI, Park SB, Son WS, Hwang DS. Midfacial soft-tissue changes after advancement of maxilla with Le Fort I osteotomy and mandibular setback surgery: comparison of conventional and high Le Fort osteotomies by superimposition of cone-beam computed tomography volumes. *J Oral Maxillofac Surg.* 2011;69:e225-e233.
73. Lloyd TE, Drage NA, Cronin AJ. The role of cone beam computed tomography in the management of unfavourable fractures following sagittal split mandibular osteotomy. *J Orthod.* 2011;38:48-54.
74. Kim YI, Choi YK, Park SB, Son WS, Kim SS. Three-dimensional analysis of dental decompensation for skeletal Class III malocclusion on the basis of vertical skeletal patterns obtained using cone-beam computed tomography. *Korean J Orthod.* 2012;42:227-234.
75. King KS, Lam EW, Faulkner MG, Heo G, Major PW. Vertical bone volume in the paramedian palate of adolescents: a computed tomography study. *Am J Orthod Dentofacial Orthop.* 2007;132:783-788.
76. Miner RM, Al Qabandi S, Rigali PH, Will LA. Cone-beam computed tomography transverse analysis. Part I: normative data. *Am J Orthod Dentofacial Orthop.* 2012;142:300-307.
77. Hilgers ML, Scarfe WC, Scheetz JP, Farman AG. Accuracy of linear temporomandibular joint measurements with cone beam computed tomography and digital cephalometric radiography. *Am J Orthod Dentofacial Orthop.* 2005;128:803-811.
78. Helenius LM, Hallikainen D, Helenius I, et al. Clinical and radiographic findings of the temporomandibular joint in patients with various rheumatic diseases: a case control study. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2005;99:455-463.
79. Bryndahl F, Eriksson L, Legrell PE, Isberg A. Bilateral TMJ disk displacement induces mandibular retrognathia. *J Dent Res.* 2006;85:1118-1123.
80. Honey OB, Scarfe WC, Hilgers MJ, et al. Accuracy of cone-beam computed tomography imaging of the temporomandibular joint: comparisons with panoramic radiology and linear tomography. *Am J Orthod Dentofacial Orthop.* 2007;132:429-438.
81. Koyama J, Nishiyama H, Hayashi T. Follow-up study of condylar bony changes using helical computed tomography in patients with temporomandibular disorder. *Dentomaxillofac Radiol.* 2007;36:472-477.
82. Ahmad M, Hollender L, Anderson Q, et al. Research diagnostic criteria for temporomandibular disorders (RDC/TMD): development of image analysis criteria and examiner reliability for image analysis. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2009;107:844-860.
83. Alexiou K, Stamatakis H, Tsiklakis K. Evaluation of the severity of temporomandibular joint osteoarthritic changes related to age using cone beam computed tomography. *Dentomaxillofac Radiol.* 2009;38:141-147.
84. Schiffman EL, Ohrbach R, Truelove EL, et al. The research diagnostic criteria for temporomandibular disorders. V: methods used to establish and validate revised Axis I diagnostic algorithms. *J Orofac Pain.* 2010;24:63-78.
85. Schiffman EL, Truelove EL, Ohrbach R, et al. The research diagnostic criteria for temporomandibular disorders. I: overview and methodology for assessment of validity. *J Orofac Pain.* 2010;24:7-24.
86. Truelove E, Pan W, Look JO, et al. The research diagnostic criteria for temporomandibular disorders. III: validity of axis I diagnoses. *J Orofac Pain.* 2010;24:35-47.
87. Alkhader M, Kuribayashi A, Ohbayashi N, Nakamura S, Kurabayashi T. Usefulness of cone beam computed tomography in temporomandibular joints with soft tissue pathology. *Dentomaxillofac Radiol.* 2010;39:343-348.
88. Tsiklakis K. Cone beam computed tomographic findings in temporomandibular joint disorders. *Alpha Omegan.* 2010;103:68-78.
89. Cevidanes LH, Hajati AK, Paniagua B, et al. Quantification of condylar resorption in temporomandibular joint osteoarthritis. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2010;110:110-117.
90. Librizzi ZT, Tadinada AS, Valiyaparambil JV, Lurie AG, Mallya SM. Cone-beam computed tomography to detect erosions of the temporomandibular joint: effect of field of view and voxel size on diagnostic efficacy and effective dose. *Am J Orthod Dentofacial Orthop.* 2011;140:e25-e30.

91. Nah KS. Condylar bony changes in patients with temporomandibular disorders: a CBCT study. *Imaging Sci Dent.* 2012;42:249-253.
92. Ferraz AM Jr, Devito KL, Guimarães JP. Temporomandibular disorder in patients with juvenile idiopathic arthritis: clinical evaluation and correlation with the findings of cone beam computed tomography. *Oral Surg Oral Med Oral Pathol Oral Radiol.* 2012;114:e51-e57.
93. Liu MQ, Chen HM, Yap AU, Fu KY. Condylar remodeling accompanying splint therapy: a cone-beam computerized tomography study of patients with temporomandibular joint disk displacement. *Oral Surg Oral Med Oral Pathol Oral Radiol.* 2012;114:259-265.
94. Barghan S, Tetradis S, Mallya S. Application of cone beam computed tomography for assessment of the temporomandibular joints. *Aust Dent J.* 2012;57(suppl 1):109-118.
95. Zain-Alabdeen EH, Alsdhan RI. A comparative study of accuracy of detection of surface osseous changes in the temporomandibular joint using multidetector CT and cone beam CT. *Dentomaxillofac Radiol.* 2012;41:185-191.
96. Falconet G, Ludlow JB, Tyndall DA, Lim PF. Correlating cone beam CT results with temporomandibular joint pain of osteoarthritic origin. *Dentomaxillofac Radiol.* 2012;41:126-130.
97. Swennen GR, Mollemans W, De Clercq C, et al. A cone-beam computed tomography triple scan procedure to obtain a three-dimensional augmented virtual skull model appropriate for orthognathic surgery planning. *J Craniofac Surg.* 2009;20:297-307.
98. Schendel SA, Lane C. 3D orthognathic surgery simulation using image fusion. *Semin Orthod.* 2009;15:48-56.
99. Ebner FH, Kürschner V, Dietz K, Bültmann E, Nägele T, Honegger J. Craniometric changes in patients with acromegaly from a surgical perspective. *Neurosurg Focus.* 2010;29:E3.
100. Edwards SP. Computer-assisted craniomaxillofacial surgery. *Oral Maxillofac Surg Clin North Am.* 2010;22:117-134.
101. Jayaratne YS, Zwahlen RA, Lo J, Tam SC, Cheung LK. Computer-aided maxillofacial surgery: an update. *Surg Innov.* 2010;17:217-225.
102. Jayaratne YS, Zwahlen RA, Lo J, Cheung LK. Three-dimensional color maps: a novel tool for assessing craniofacial changes. *Surg Innov.* 2010;17:198-205.
103. Popat H, Richmond S. New developments in: three-dimensional planning for orthognathic surgery. *J Orthod.* 2010;37:62-71.
104. Carvalho Fde A, Cevidanes LH, da Motta AT, Almeida MA, Phillips C. Three-dimensional assessment of mandibular advancement 1 year after surgery. *Am J Orthod Dentofacial Orthop.* 2010;137(4 suppl):S53.e1-S53.e12.
105. da Motta AT, de Assis Ribeiro Carvalho F, Oliveira AE, Cevidanes LH, de Oliveira Almeida MA. Superimposition of 3D cone-beam CT models in orthognathic surgery. *Dent Press J Orthod.* 2010;15:39-41.
106. Agarwal R. Anthropometric evaluation of complete unilateral cleft lip nose with cone beam CT in early childhood. *J Plast Reconstr Aesthet Surg.* 2011;64:e181-e182.
107. Behnia H, Khojasteh A, Soleimani M, Tehranchi A, Atashi A. Repair of alveolar cleft defect with mesenchymal stem cells and platelet derived growth factors: a preliminary report. *J Cranio-maxillofac Surg.* 2012;40:2-7.
108. Dalessandri D, Laffranchi L, Tonni I, et al. Advantages of cone beam computed tomography (CBCT) in the orthodontic treatment planning of cleidocranial dysplasia patients: a case report. *Head Face Med.* 2011;7:6.
109. Abou-Elfetouh A, Barakat A, Abdel-Ghany K. Computed-guided rapid-prototyped templates for segmental mandibular osteotomies: a preliminary report. *Int J Med Robot.* 2011;7:187-192.
110. Scolozzi P, Terzic A. "Mirroring" computational planning, navigation guidance system, and intraoperative mobile C-arm cone-beam computed tomography with flat-panel detector. *J Oral Maxillofac Surg.* 2011;69:1697-1707.
111. Aboudara CA, Hatcher D, Nielsen IL, Miller A. A three-dimensional evaluation of the upper airway in adolescents. *Orthod Craniofac Res.* 2003;6(suppl 1):173-175.
112. Sera T, Fujioka H, Yokota H, et al. Three-dimensional visualization and morphometry of small airways from microfocal X-ray computed tomography. *J Biomech.* 2003;36:1587-1594.
113. Ogawa T, Enciso R, Memon A, Mah JK, Clark GT. Evaluation of 3D airway imaging of obstructive sleep apnea with cone-beam computed tomography. *Stud Health Technol Inform.* 2005;111:365-368.
114. Strauss RA, Burgoyne CC. Diagnostic imaging and sleep medicine. *Dent Clin North Am.* 2008;52:891-915.
115. Osorio F, Perilla M, Doyle DJ, Palomo JM. Cone beam computed tomography: an innovative tool for airway assessment. *Anesth Analg.* 2008;106:1803-1807.
116. Tso HH, Lee JS, Huang JC, Maki K, Hatcher D, Miller AJ. Evaluation of the human airway using cone-beam computerized tomography. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2009;108:768-776.
117. Lenza MG, Lenza MM, Dalstra M, Melsen B, Cattaneo PM. An analysis of different approaches to the assessment of upper airway morphology: a CBCT study. *Orthod Craniofac Res.* 2010;13:96-105.
118. El H, Palomo JM. Measuring the airway in 3 dimensions: a reliability and accuracy study. *Am J Orthod Dentofacial Orthop.* 2010;137(4 suppl):S50.e1-S50.e9.
119. Schendel SA, Hatcher D. Automated 3-dimensional airway analysis from cone-beam computed tomography data. *J Oral Maxillofac Surg.* 2010;68:696-701.
120. El AS, El H, Palomo JM, Baur DA. A 3-dimensional airway analysis of an obstructive sleep apnea surgical correction with cone beam computed tomography. *J Oral Maxillofac Surg.* 2011;69:2424-2436.
121. Oh KM, Hong JS, Kim YJ, Cevidanes LS, Park YH. Three-dimensional analysis of pharyngeal airway form in children with anteroposterior facial patterns. *Angle Orthod.* 2011;81:1075-1082.
122. Abramson Z, Susarla SM, Lawler M, Bouchard C, Troulis M, Kaban LB. Three-dimensional computed tomographic airway analysis of patients with obstructive sleep apnea treated by maxillomandibular advancement. *J Oral Maxillofac Surg.* 2011;69:677-686.
123. Schendel S, Powell N, Jacobson R. Maxillary, mandibular, and chin advancement: treatment planning based on airway anatomy in obstructive sleep apnea. *J Oral Maxillofac Surg.* 2011;69:663-676.
124. Iwasaki T, Saitoh I, Takemoto Y, et al. Evaluation of upper airway obstruction in Class II children with fluid-mechanical simulation. *Am J Orthod Dentofacial Orthop.* 2011;139:e135-e145.
125. Conley RS. Evidence for dental and dental specialty treatment of obstructive sleep apnoea. Part 1: the adult OSA patient and Part 2: the paediatric and adolescent patient. *J Oral Rehabil.* 2011;38:136-156.
126. de Souza Carvalho AC, Magro Filho O, Garcia IR Jr, Araujo PM, Nogueira RL. Cephalometric and three-dimensional assessment of superior posterior airway space after maxillomandibular advancement. *Int J Oral Maxillofac Surg.* 2012;41:1102-1111.
127. Lee Y, Chun YS, Kang N, Kim M. Volumetric changes in the upper airway after bimaxillary surgery for skeletal Class III

- malocclusions: a case series study using 3-dimensional cone-beam computed tomography. *J Oral Maxillofac Surg.* 2012;70:2867-2875.
128. Farronato G, Storti E, Cuzzocrea ML, et al. Three-dimensional changes of the upper airway in patients with obstructive sleep apnea syndrome after a non-adjustable oral appliance treatment. *Minerva Stomatol.* 2013;62:107-116.
 129. Raffaini M, Pisani C. Clinical and cone-beam computed tomography evaluation of the three-dimensional increase in pharyngeal airway space following maxillo-mandibular rotation-advancement for Class II-correction in patients without sleep apnoea (OSA). *J Craniomaxillofac Surg*; 2013. <http://dx.doi.org/10.1016/j.jcms.2012.11.022> [e-pub ahead of print].
 130. Kim MA, Kim BR, Choi JY, Youn JK, Kim YJ, Park YH. Three-dimensional changes of the hyoid bone and airway volumes related to its relationship with horizontal anatomic planes after bimaxillary surgery in skeletal Class III patients. *Angle Orthod.* 2013;83:623-629.
 131. Iwasaki T, Saitoh I, Takemoto Y, et al. Tongue posture improvement and pharyngeal airway enlargement as secondary effects of rapid maxillary expansion: a cone-beam computed tomography study. *Am J Orthod Dentofacial Orthop.* 2013;143:235-245.
 132. Weissheimer A, Menezes LM, Sameshima GT, Enciso R, Pham J, Grauer D. Imaging software accuracy for 3-dimensional analysis of the upper airway. *Am J Orthod Dentofacial Orthop.* 2012;142:801-813.
 133. Alsufyani NA, Flores-Mir C, Major PW. Three-dimensional segmentation of the upper airway using cone beam CT: a systematic review. *Dentomaxillofac Radiol.* 2012;41:276-284.
 134. Poggio PM, Incorvati C, Velo S, Carano A. "Safe zones": a guide for miniscrew positioning in the maxillary and mandibular arch. *Angle Orthod.* 2006;76:191-197.
 135. Gracco A, Lombardo L, Cozzani M, Siciliani G. Quantitative evaluation with CBCT of palatal bone thickness in growing patients. *Prog Orthod.* 2006;7:164-174.
 136. King KS, Lam EW, Faulkner MG, Heo G, Major PW. Predictive factors of vertical bone depth in the paramedian palate of adolescents. *Angle Orthod.* 2006;76:745-751.
 137. Palomo L, Palomo JM, Hans MG, Bissada N. Image guided placement of temporary anchorage devices for tooth movement. *Int J Comput Assist Radiol Surg.* 2007;2(suppl 1):S424-S426.
 138. Gracco A, Luca L, Cozzani M, Siciliani G. Assessment of palatal bone thickness in adults with cone beam computerised tomography. *Aust Orthod J.* 2007;23:109-113.
 139. Ono A, Motoyoshi M, Shimizu N. Cortical bone thickness in the buccal posterior region for orthodontic mini-implants. *Int J Oral Maxillofac Surg.* 2008;37:334-340.
 140. Gracco A, Lombardo L, Cozzani M, Siciliani G. Quantitative cone-beam computed tomography evaluation of palatal bone thickness for orthodontic miniscrew placement. *Am J Orthod Dentofacial Orthop.* 2008;134:361-369.
 141. Kim GT, Kim SH, Choi YS, et al. Cone-beam computed tomography evaluation of orthodontic miniplate anchoring screws in the posterior maxilla. *Am J Orthod Dentofacial Orthop.* 2009;136:628.e1-628.e10.
 142. Park J, Cho HJ. Three-dimensional evaluation of interradicular spaces and cortical bone thickness for the placement and initial stability of microimplants in adults. *Am J Orthod Dentofacial Orthop.* 2009;136:314.e1-314.e12.
 143. Kim SH, Yoon HG, Choi YS, Hwang EH, Kook YA, Nelson G. Evaluation of interdental space of the maxillary posterior area for orthodontic mini-implants with cone-beam computed tomography. *Am J Orthod Dentofacial Orthop.* 2009;135:635-641.
 144. Baumgaertel S. Quantitative investigation of palatal bone depth and cortical bone thickness for mini-implant placement in adults. *Am J Orthod Dentofacial Orthop.* 2009;136:104-108.
 145. Baumgaertel S, Hans MG. Buccal cortical bone thickness for mini-implant placement. *Am J Orthod Dentofacial Orthop.* 2009;136:230-235.
 146. Baumgaertel S, Hans MG. Assessment of infrazygomatic bone depth for mini-screw insertion. *Clin Oral Implants Res.* 2009;20:638-642.
 147. Kau CH, English JD, Muller-Delgado MG, Hamid H, Ellis RK, Winklemann S. Retrospective cone-beam computed tomography evaluation of temporary anchorage devices. *Am J Orthod Dentofacial Orthop.* 2010;137:166.e1-166.e5.
 148. Park HS, Hwangbo ES, Kwon TG. Proper mesiodistal angles for microimplant placement assessed with 3-dimensional computed tomography images. *Am J Orthod Dentofacial Orthop.* 2010;137:200-206.
 149. Fayed MM, Pazera P, Katsaros C. Optimal sites for orthodontic mini-implant placement assessed by cone beam computed tomography. *Angle Orthod.* 2010;80:939-951.
 150. Morea C, Hayek JE, Oleskovicz C, Dominguez GC, Chilvarquer I. Precise insertion of orthodontic miniscrews with a stereolithographic surgical guide based on cone beam computed tomography data: a pilot study. *Int J Oral Maxillofac Implants.* 2011;26:860-865.
 151. Qiu L, Haruyama N, Suzuki S, et al. Accuracy of orthodontic miniscrew implantation guided by stereolithographic surgical stent based on cone-beam CT-derived 3D images. *Angle Orthod.* 2012;82:284-293.
 152. Garrett BJ, Caruso JM, Rungcharassaeng K, Farrage JR, Kim JS, Taylor GD. Skeletal effects to the maxilla after rapid maxillary expansion assessed with cone-beam computed tomography. *Am J Orthod Dentofacial Orthop.* 2008;134:8-9.
 153. Christie KF, Boucher N, Chung CH. Effects of bonded rapid palatal expansion on the transverse dimensions of the maxilla: a cone-beam computed tomography study. *Am J Orthod Dentofacial Orthop.* 2010;137(4 suppl):S79-S85.
 154. Tai K, Park JH. Dental and skeletal changes in the upper and lower jaws after treatment with Schwarz appliances using cone-beam computed tomography. *J Clin Pediatr Dent.* 2010;35:111-120.
 155. Domann CE, Kau CH, English JD, Xia JJ, Souccar NM, Lee RP. Cone beam computed tomography analysis of dentoalveolar changes immediately after maxillary expansion. *Orthod (Chic).* 2011;12:202-209.
 156. Baratieri C, Alves M Jr, Sant'anna EF, Nojima Mda C, Nojima LI. 3D mandibular positioning after rapid maxillary expansion in Class II malocclusion. *Braz Dent J.* 2011;22:428-434.
 157. Baysal A, Karadede I, Hekimoglu S, et al. Evaluation of root resorption following rapid maxillary expansion using cone-beam computed tomography. *Angle Orthod.* 2012;82:488-494.
 158. Tai K, Park JH, Mishima K, Shin JW. 3-Dimensional cone-beam computed tomography analysis of transverse changes with Schwarz appliances on both jaws. *Angle Orthod.* 2011;81:670-677.
 159. Silva MA, Wolf U, Heinicke F, Bumann A, Visser H, Hirsch E. Cone-beam computed tomography for routine orthodontic treatment planning: a radiation dose evaluation. *Am J Orthod Dentofacial Orthop.* 2008;133:640.e1-640.e5.
 160. Larson BE. Cone-beam computed tomography is the imaging technique of choice for comprehensive orthodontic assessment. *Am J Orthod Dentofacial Orthop.* 2012;141:402, 404, 406 passim.
 161. Hodges RJ, Atchison KA, White SC. Impact of cone-beam computed tomography on orthodontic diagnosis and treatment

- planning. *Am J Orthod Dentofacial Orthop.* 2013;143:665-674.
162. van Vlijmen OJ, Kuijpers MA, Bergé SJ, et al. Evidence supporting the use of cone-beam computed tomography in orthodontics. *J Am Dent Assoc.* 2012;143:241-252.
 163. Fryback DG, Thornbury JR. The efficacy of diagnostic imaging. *Med Decis Mak.* 1991;11:88-94.
 164. National Health and Medical Research Council of Australia. *A Guide to the Development, Implementation and Evaluation of Clinical Practice Guidelines*; 1999. Available at: http://www.nhmrc.gov.au/_files_nhmrc/publications/attachments/cp30.pdf. Accessed September 23, 2011.
 165. Cascade PN. The American College of Radiology. ACR Appropriateness Criteria project. *Radiology.* 2000;214(suppl):3-46.
 166. U.S. Preventive Services Task Force Ratings: Grade Definitions. *Guide to Clinical Preventive Services, Third Edition: Periodic Updates*; 2000-2003. Available at: <http://www.uspreventiveservicestaskforce.org/3rduspstf/ratings.htm>. Accessed September 23, 2011.
 167. European Commission. Radiation Protection 136. European Guidelines on Radiation Protection in Dental Radiology. 2004: 115. ISBN 92-894-5958-1
 168. American College of Radiology. *ACR Appropriateness Criteria. Rating Round Information*; 2011. Available at: http://www.acr.org/SecondaryMainMenuCategories/quality_safety/app_criteria/Rating-Round-Information.aspx. Accessed September 23, 2011.
 169. Claus EB, Calvocoressi L, Bondy ML, Schildkraut JM, Wiemels JL, Wrensch M. Dental X-rays and risk of meningioma. *Cancer*; 2012. <http://dx.doi.org/10.1002/encr.26625>. Accessed June 7, 2012.
 170. Jorgensen TJ. Dental X-rays and risk of meningioma. *Cancer.* 2013;119:463.
 171. Calnon WR. Shortcomings of study on dental X-rays and risk of meningioma. *Cancer.* 2013;119:464-465.
 172. Tetradis S, White SC, Service SK. Dental x-rays and risk of meningioma; the jury is still out. *J Evid Based Dent Pract.* 2012;12:174-177.
 173. Dirksen D, Runte C, Berghoff L, Scheutzel P, Figgenger L. Dental X-rays and risk of meningioma: anatomy of a case-control study. *J Dent Res.* 2013;92:397-398.
 174. Pearce MS, Salotti JA, Little MP, et al. Radiation exposure from CT scans in childhood and subsequent risk of leukaemia and brain tumours: a retrospective cohort study. *Lancet.* 2012;380:499-505.
 175. Smith-Bindman R, Lipson J, Marcus R, et al. Radiation dose associated with common computed tomography examinations and the associated lifetime attributable risk of cancer. *Arch Intern Med.* 2009;169:2078-2086.
 176. Ludlow JB, Davies-Ludlow LE, Brooks SL, Howerton WB. Dosimetry of 3 CBCT devices for oral and maxillofacial radiology: CB Mercuray, NewTom 3G and i-CAT. *Dentomaxillofac Radiol.* 2006;35:219-226.
 177. Ludlow JB, Davies-Ludlow LE, White SC. Patient risk related to common dental radiographic examinations: the impact of 2007 International Commission on Radiological Protection recommendations regarding dose calculation. *J Am Dent Assoc.* 2008;139:1237-1243.
 178. Loubele M, Bogaerts R, Van Dijk E, et al. Comparison between effective radiation dose of CBCT and MSCT scanners for dentomaxillofacial applications. *Eur J Radiol.* 2009;71:461-468.
 179. Suomalainen A, Kiljunen T, Käser Y, Peltola J, Kortensniemi M. Dosimetry and image quality of four dental cone beam computed tomography scanners compared with multislice computed tomography scanners. *Dentomaxillofac Radiol.* 2009;38:367-378.
 180. Okano T, Harata Y, Sugihara Y, et al. Absorbed and effective doses from cone beam volumetric imaging for implant planning. *Dentomaxillofac Radiol.* 2009;38:79-85.
 181. Carrafiello G, Dizonno M, Colli V, et al. Comparative study of jaws with multislice computed tomography and cone-beam computed tomography. *Radiol Med.* 2010;115:600-611.
 182. Jeong DK, Lee SC, Huh KH, et al. Comparison of effective dose for imaging of mandible between multi-detector CT and cone-beam CT. *Imaging Sci Dent.* 2012;42:65-70.
 183. Theodorakou C, Walker A, Horner K, Pauwels R, Bogaerts R, Jacobs R; SEDENTEXCT Project Consortium. Estimation of paediatric organ and effective doses from dental cone beam CT using anthropomorphic phantoms. *Br J Radiol.* 2012;85:153-160.
 184. Ludlow JB, Ivanovic M. Comparative dosimetry of dental CBCT devices and 64-slice CT for oral and maxillofacial radiology. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2008;106:106-114.
 185. Hirsch E, Wolf U, Heinicke F, Silva MA. Dosimetry of the cone beam computed tomography Veraviewepocs 3D compared with the 3D Accuitomo in different fields of view. *Dentomaxillofac Radiol.* 2008;37:268-273.
 186. Roberts JA, Drage NA, Davies J, Thomas DW. Effective dose from cone beam CT examinations in dentistry. *Br J Radiol.* 2009;82:35-40.
 187. Qu XM, Li G, Ludlow JB, Zhang ZY, Ma XC. Effective radiation dose of ProMax 3D cone-beam computerized tomography scanner with different dental protocols. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2010;110:770-776.
 188. Ludlow JB. A manufacturer's role in reducing the dose of cone beam computed tomography examinations: effect of beam filtration. *Dentomaxillofac Radiol.* 2011;40:115-122.
 189. Lofthag-Hansen S, Thilander-Klang A, Gröndahl K. Evaluation of subjective image quality in relation to diagnostic task for cone beam computed tomography with different fields of view. *Eur J Radiol.* 2011;80:483-488.
 190. Davies J, Johnson B, Drage NA. Effective doses from cone beam CT investigation of the jaws. *Dentomaxillofac Radiol.* 2012;41:30-36.
 191. Pauwels R, Beinsberger J, Collaert B, et al; The SEDENTEXCT Project Consortium. Effective dose range for dental cone beam computed tomography scanners. *Eur J Radiol.* 2012;81:267-271.
 192. Grünheid T, Kolbeck Schieck JR, Pliska BT, Ahmad M, Larson BE. Dosimetry of a cone-beam computed tomography machine compared with a digital X-ray machine in orthodontic imaging. *Am J Orthod Dentofacial Orthop.* 2012;141:436-443.
 193. Qu XM, Li G, Sanderink GC, Zhang ZY, Ma XC. Dose reduction of cone beam CT scanning for the entire oral and maxillofacial regions with thyroid collars. *Dentomaxillofac Radiol.* 2012;41:373-378.
 194. Koivisto J, Kiljunen T, Tapiovaara M, Wolff J, Kortensniemi M. Assessment of radiation exposure in dental cone-beam computerized tomography with the use of metal-oxide semiconductor field-effect transistor (MOSFET) dosimeters and Monte Carlo simulations. *Oral Surg Oral Med Oral Pathol Oral Radiol.* 2012;114:393-400.
 195. Palomo JM, Rao PS, Hans MG. Influence of CBCT exposure conditions on radiation dose. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod.* 2008;105:773-782.
 196. Schilling R, Geibel MA. Assessment of the effective doses from two dental cone beam CT devices. *Dentomaxillofac Radiol.* 2013;42:20120273.

197. Valentin J. The 2007 recommendations of the International Commission on Radiological Protection. Publication 93. *Ann ICRP*. 2007;37:1-332.
198. Preston DL, Shimizu Y, Pierce DA, Suyama A, Mabuchi K. Studies of mortality of atomic bomb survivors. Report 13: solid cancer and non-cancer disease mortality: 1950-1997. *Radiat Res*. 2003;160:381-407.
199. United Nations Scientific Committee on the Effects of Atomic Radiation. *Effects of Ionizing Radiation: United Nations Scientific Committee on the Effects of Atomic Radiation – UNSCEAR 2006 Report, Volume 1-Report to the General Assembly, With Scientific Annexes A and B*. New York, NY: United Nations; 2008:360.
200. National Research Council (U.S.), Committee to Assess Health Risks from Exposure to Low Levels of Ionizing Radiation. *Health Risks From Exposure to Low Levels of Ionizing Radiation: BEIR VII – Phase 2*. Washington, DC: The National Academies Press, ISBN: 0-309-53040-7. 2006:424.
201. International Commission on Radiological Protection, 1990 recommendations of the International Commission on Radiological Protection, ICRP publication 60. *Ann ICRP*. 1991;21:1-3.
202. Brenner DJ, Elliston CD, Hall EJ, Berdon WE. Estimated risks of radiation-induced fatal cancer from pediatric CT. *Am J Roentgenol*. 2001;176:289-296.
203. American College of Radiology. *ACR Appropriateness Criteria. Radiation Dose Assessment Introduction*; 2011. Available at: <http://www.acr.org/~media/ACR/Documents/AppCriteria/RRLLInformation.pdf>. Accessed May 22, 2012.
204. The Alliance for Radiation Safety in Pediatric Imaging. *Image Gently*; 2011. Available at: <http://www.imagegently.org>. Accessed December 21, 2011.
205. American College of Radiology. *The Quality Improvement Registry for CT Scans in Children*; 2010. Available at: <https://nrdr.acr.org/Portal/QuIRCC/Main/page.aspx>. Accessed December 21, 2011.
206. Kwong JC, Palomo JM, Landers MA, Figueroa A, Hans MG. Image quality produced by different CBCT settings. *Am J Orthod Dentofacial Orthop*. 2008;133:317-327.
207. Gavala S, Donta C, Tsiklakis K, Boziari A, Kamenopoulou V, Stamatakis HC. Radiation dose reduction in direct digital panoramic radiography. *Eur J Radiol*. 2009;71:42-48.
208. National Research Council of the National Academies, Committee to Assess Health Risks from Exposure to Low Levels of Ionizing Radiation. *Health Risks From Exposure to Low Levels of Ionizing Radiation: BEIR VII – Phase 2*. Washington, DC: The National Academies Press; ISBN: 0-309-53040-7, 2006:424.
209. Cha JY, Mah J, Sinclair P. Incidental findings in the maxillofacial area with 3 dimensional cone beam imaging. *Am J Orthod Dentofacial Orthop*. 2007;132:7-14.
210. Pliska B, DeRocher M, Larson BE. Incidence of significant findings on CBCT scans of an orthodontic patient population. *Northwest Dent*. 2011;90:12-16.
211. Pazera P, Bornstein MM, Pazera A, Sendi P, Katsaros C. Incidental maxillary sinus findings in orthodontic patients: a radiographic analysis using cone-beam computed tomography (CBCT). *Orthod Craniofac Res*. 2011;14:17-24.
212. Gracco A, Incerti Parenti S, Ioele C, Alessandri Bonetti G, Stellini E. Prevalence of incidental maxillary sinus findings in Italian orthodontic patients: a retrospective cone-beam computed tomography study. *Korean J Orthod*. 2012;42:329-334.
213. Drage N, Rogers S, Greenall C, Playle R. Incidental findings on cone beam computed tomography in orthodontic patients. *J Orthod*. 2013;40:29-37.
214. Rogers SA, Drage N, Durning P. Incidental findings arising with cone beam computed tomography imaging of the orthodontic patient. *Angle Orthod*. 2011;81:350-355.
215. Carter L, Farman A, Geist J, et al. American Academy of Oral and Maxillofacial Radiology executive opinion statement on performing and interpreting diagnostic cone beam computed tomography. *Oral Surg Oral Med Oral Path Oral Radiol Endod*. 2008;106:561-562.
216. Bowlin J. Cone beam technology: legal caveats. *Bull Am Assoc Orthodontists*. 2010;28:24-25.

APPENDIX A: DIAGNOSTIC USES OF CBCT IN ORTHODONTICS

Dental structural anomalies

These comprise assessments of variations in tooth morphology, hypodontia, retained primary teeth, supernumeraries/generation/fusion, root abnormalities, and external and internal resorption.²¹⁻³²

Anomalies in dental position

These include dental impactions, presence of unerupted and impacted supernumeraries, locations of molars in relation to the inferior alveolar canals, anomalies in eruption sequences, and ectopic eruptions (including teeth in clefts).^{14,22,27,33-50}

Compromised dento-alveolar boundaries

The assessment of dento-alveolar volume (in addition to that which can be determined by clinical examination and study models) is needed when there is reduced buccal/lingual alveolar width, bimaxillary protrusion, compromised periodontal status, and/or clefts of the alveolus.⁵¹⁻⁵⁶

Asymmetry

Clinically, asymmetry presents as chin or mandibular deviation, dental midline deviation, and/or occlusal cant discrepancies as well as other dental and craniofacial asymmetries.⁵⁷⁻⁶⁴

Anteroposterior discrepancies

These are skeletally based Class II and Class III malocclusions.^{59,60,62,65-73}

Vertical discrepancies

Initial facial patterns assessed clinically or radiographically may suggest skeletal discrepancies related to vertical maxillary deficiency or excess and may present as anterior open bite or deep overbite.^{67,74}

Transverse discrepancies

These anomalies may be present as either skeletal lingual or buccal crossbites or discrepancies without the

presence of crossbites in which there is excessive dental compensation of the bucco-lingual inclination of posterior teeth.^{67,75,76}

Temporomandibular joint (TMJ) signs and/or symptoms

TMJ pathoses that result in alterations in the size, form, quality and spatial relationships of the osseous joint components may lead to skeletal and dental discrepancies in the three planes of space. In affected condyles, perturbed resorption and/or apposition can lead to progressive bite changes and compensations in the maxilla. In addition, tooth position, occlusion and the articular fossa of the non-affected side of the mandible can become involved. The sequelae of these changes are unpredictable orthodontic outcomes. Such TMJ conditions include developmental disorders such as condylar hyperplasia, hypoplasia, or aplasia, arthritic degeneration, persistently symptomatic joints, and bite changes including progressive bite opening and limitation or deviation upon opening or closing.⁷⁷⁻⁹⁶

Dentofacial deformities and craniofacial anomalies

CBCT imaging can facilitate analysis of these conditions and be used to simulate virtual treatments and plan orthopedic corrections and orthognathic surgeries. Computer-aided jaw surgery is increasing in use clinically because virtual plans accurately represent surgical procedures in the operating room.^{65,66,68-73,97-109}

Conditions that affect airway morphology

A number of authors have used CBCT imaging to measure airway dimensions and reported changes over time with specific therapies including orthognathic surgery and particularly obstructive sleep apnea.^{18,111-131} There are challenges in the use of CBCT clinically as the validity of such measurements may vary.^{132,133} The boundaries of the nasopharynx with the maxillary/paranasal sinuses and of the oropharynx with the oral cavity are often not consistent among subjects and image acquisitions, and airway shapes and volumes vary markedly with dynamic processes such as breathing and head postures.

In addition, CBCT has been reported useful in preoperative assessment and/or postoperative evaluation of treatment outcomes for specific research applications including:

Specific surgical procedures

Research in the areas of craniofacial growth and development as well as assessments of the short- and long-term outcomes of various treatment regimens has the potential to benefit from CBCT assessments of longitudinal changes and diagnostic characterization of

tooth and facial morphology of hard and soft tissues. Studies on the morphological basis for craniofacial growth and response to treatment can help elucidate clinical questions on variability of outcomes of treatment, as well as clarify treatment effects and areas of bone remodeling and displacement.

Orthodontic mini-implants used as temporary anchorage devices

Numerous authors have identified CBCT imaging as being clinically useful in identifying optimal site location for placement of orthodontic mini-implants.^{67,75,134-151}

Maxillary expanders

CBCT imaging of maxillary transverse deficiencies treated with fixed and removable expanders has been reported of benefit in characterizing appliance specific skeletal displacement, associated dental effects and quantifying changes in skeletal dimensions of the nasal cavity and maxillary sinus volume.^{51,152-158}

APPENDIX B: RATIONAL FOR ORTHODONTIC IMAGE SELECTION RECOMMENDATIONS

The recommendations in Table III are based upon the complexity of the orthodontic case. The following were considered in developing the recommendations.

Selection of clinical conditions for indications of CBCT use

The most common clinical dental and skeletal conditions in the orthodontic patient are presented as column headings in Table III.

Definition of orthodontic treatment difficulty criteria

The panel acknowledges the uniqueness of the facial form of each patient and the inherent difficulty in attempting to assess the severity of malocclusion and quantifying and categorizing orthodontic treatment need. For patients with severe malocclusions, there are, however, more choices with regard to appropriate orthodontic treatments, and there is an increased need for radiographic diagnostic input. For Table III, malocclusion severity was categorized and anticipated appropriateness of CBCT imaging was listed according to three levels of patient presentation:

Mild. Patients present with dental malocclusions, with or without minimal anteroposterior, vertical, or transverse skeletal discrepancies. These patients are treated usually with conventional biomechanics (with or without extraction). CBCT imaging is likely

inappropriate for these patients unless they present with the additional clinical conditions noted.

Moderate. Patients present with dental and skeletal discrepancies that are treated orthodontically and/or orthopedically only. These discrepancies include bimaxillary proclination, open bite, and compensated Class III malocclusion. CBCT imaging is possibly indicated for many of these patients as indicated.

Severe. Patients present with skeletal conditions including, but not limited to complicated skeletal discrepancies, craniofacial anomalies (e.g., cleft lip and palate, craniofacial synostosis, etc.), sleep apnea, speech disorders, and post oncology/trauma/resection/pathology. For patients in this group, a team approach for treatment is used including speech therapy, clinical psychology, orthodontic and surgical interventions. Advanced imaging, including CBCT, may be indicated for many of these patients.

Selection of FOV

There is limited published research on the many and varied technical issues associated with CBCT imaging in orthodontics including optimal fields of view (image sizes) for specific diagnostic tasks, optimal exposure settings (some tasks may require lower exposures than others), and variations in the levels of ionizing radiation used (for similar tasks) with various CBCT systems. More specific and additional issues and controversies related to CBCT use include: 1) the necessary diagnostic quality of images²⁰⁵; 2) imperfect superimposition of CBCT and surface-scan data; 3) differing levels of exposure needed to determine root and bone morphology related to appliance construction or for the diagnosis of pathology; 4) indications for use of multiple CBCT scans; 5) lack of and utility of 3D norms; 6) impact of CBCT for the assessment of treatment outcome; 7) responsibility for the identification of clinically significant incidental pathology; and 8) responsibility for calibration and maintenance of the equipment.²⁰³

Assessment of progress and treatment outcomes

In complex cases, follow-up CBCT acquisitions for growth observation, assessment of treatment progress, and posttreatment analysis may be helpful. Any imaging protocol for the longitudinal quantitative assessment of the craniofacial complex requires methods to: 1) minimize the radiation dose from sequential multiple CBCT exposures; 2) construct accurate three-dimensional surface models; 3) reliably image registration (non-rigid, elastic and deformable; or rigid registration) using stable structures of reference for cranial base or regional superimpositions; and 4) quantify changes over time.

Age considerations

The choice of radiographic imaging method of a patient with clinically determined dental and/or skeletal modifying factors is dependent on the stage of growth of the individual and age-related presentation of the condition; therefore, recommendations for CBCT for some dental/skeletal conditions are age dependent. These conditions include:

Tooth structural anomalies. A CBCT examination may be indicated when other diagnostic modalities indicate a problem with root morphology or resorption in the mixed and permanent dentitions.

Tooth positional or eruption anomalies. A possible indication for a CBCT examination (in addition to periapical, occlusal and/or panoramic images) exists when interceptive orthodontic treatment is being considered for children between the ages of 5-11. In such cases, a small FOV should be used. Another possible indication for a CBCT examination (usually restricted or small FOV) is for children more than 11 years of age if surgical exposure is being considered as a treatment option and the location of the crown cannot be determined clinically or with conventional 2D images (e.g., panoramic, occlusal and/or periapical images).

Craniofacial anomalies. An additional possible indication for CBCT is in children (0-4 years) prior to mandibular distraction or other craniofacial surgical treatments if the children can remain motionless during the scans. For children between 5 and 11 years of age, CBCT is useful for locating developing teeth prior to alveolar bone grafting and Phase I orthodontic treatment for children with oral clefts. For these cases, limited fields of views may suffice. For patients older than 11 and comprehensive orthodontic treatments are required in preparation for craniofacial surgical procedures, CBCT may provide a benefit at the diagnostic stage of orthodontic treatment as well as immediately before the surgical procedures. Such decisions are case specific.

APPENDIX C: CALCULATION OF RRL FOR ORTHODONTIC IMAGING

Table VI provides four orthodontic imaging protocols and provides RRLs^{168,203} and published effective doses. For example, if a typical imaging protocol incorporates three digital (Planmeca PM Proline 2000 [low dose]) panoramic images (initial- diagnostic-, mid- and post-treatment; 12 μSv ²⁰⁷ for each exposure = 36 μSv) and two digital (photo-stimulable storage phosphor) lateral cephalometric images (initial- and post-treatment; 5.6 μSv ¹⁷⁷ for each exposure = 11.2 μSv) the total equivalent dose for the orthodontic series is 47.2 μSv . For an adult this represents an RRL of ☉ whereas for

a child this represents an RRL of ☹☹. This can be compared to orthodontic imaging series incorporating a large FOV CBCT (i-CAT Next Generation [16 × 13 cm]) image (initial; 83 μSv ¹⁹¹), two digital (Planmeca PM Proline 2000 [low dose]²⁰⁶) panoramic images (mid- and post-treatment; 12 μSv ²⁰⁷ for each exposure = 24 μSv) and one digital (photo-stimulable storage phosphor) lateral cephalometric image (post-treatment; 5.6 μSv ¹⁷⁷). The equivalent dose for this orthodontic imaging series is 112.6 μSv . While radiation risk (RRL) using CBCT in this example is for both the

adult and child is the same (☹☹), this protocol provides over twice the absolute dose than the conventional imaging series and elevates the risk of the adult into a higher category.

Reprint requests:

William C. Scarfe, BDS, MS, FRACDS
Department of Surgical and Hospital Dentistry
School of Dentistry, University of Louisville
Louisville, KY 40292, USA
william.scarfe@louisville.edu; wscar01@louisville.edu